



NORTH BAY PARRY SOUND DISTRICT HEALTH UNIT

PUBLIC HEALTH UNIT INFECTION PREVENTION AND CONTROL LAPSE REPORT	
Initial Report	
Premise/facility under investigation (name and address)	Bonnie Brown Foot Care
Type of premise/facility: (E.g. clinic, personal services setting)	Foot Care Services (home and mobile)
Date Board of Health became aware of IPAC lapse	1/18/2019
Date of Initial Report posting	1/21/2019
Date of Initial Report update(s) (if applicable)	n/a
How the IPAC lapse was identified	Other
Summary Description of the IPAC Lapse	Complaint received 1/17/2019. Critical medical equipment not sufficiently reprocessed. Furniture is can not be easily cleaned and disinfected. An approved sharps container was not present. Education related to general IPAC practices and reprocessing is not sufficient.
IPAC Lapse Investigation	
Did the IPAC lapse involve a member of a regulatory college?	Yes
If yes, was the issue referred to the regulatory college?	Yes
Were any corrective measures recommended and/or implemented?	Yes
Please provide further details/steps	Critical medical equipment is required to be cleaned and sterilized. Furniture is required to be easily cleaned and disinfected. A sharps container is required. Education related to general IPAC practices and reprocessing is required.
Date any order(s) or directive(s) were issued to the owners/operators (if applicable)	Verbal requirements on 1/18/2019, written on 1/22/2019.
Initial Report Comments and Contact Information	
Any Additional Comments (Do not include any personal information or personal health information)	Complaint investigation was conducted. Corrective measures were required as indicated above.
If you have any further questions, please contact:	
Name	Robert A-Muhong
Title	Program Manager, Environmental Health
E-mail address	robert.a-muhong@healthunit.ca
Phone number	(705) 474-1400 x5320
Final Report	
Date of Final Report posting:	2/20/2019
Date any order(s) or directive(s) were issued to the owner/operator (if applicable)	Verbal requirements on 1/18/2019, written on 1/22/2019.
Brief description of corrective measures taken	A third party sterilizing contract is in place and single use disposable instruments also available for use. Sharps container is present as required. Client furniture is easily cleaned and disinfected. Education modules completed.
Date all corrective measures were confirmed to have been completed	2/14/2019
Final Report Comments and Contact Information	
Any Additional Comments (Do not include any personal information or personal health information)	Re-inspection completed 1/31/2019; outstanding education requirements completed 2/14/2019.
If you have any further questions, please contact:	
Name	Robert A-Muhong
Title	Program Manager, Environmental Health
Email address	robert.a-muhong@healthunit.ca
Phone number	(705) 474-1400 ext. 5320